

JRT Associates
5 Nepperhan Avenue, Suite 2B
Elmsford, NY 10523
800-221-0111

**Accuracy Test Data
Dose Calibrator AMA office**

Biodex Model 086-332 Serial # 08070015 Detector Serial # 5300120

Date	Sealed Source	Reading				Calculated Standard	Acceptable Variance +/- 5%	Technologist
		1	2	3	AVG			
04/15/09	Cs-137	208.	207.	208.	208.	209. μ Ci	-0.6%	AMA
04/15/09	Ba-133	248.	248.	248.	248.	255. μ Ci	-2.8%	AMA
04/15/09	Ce-57	1589.	1581.	1588.	1586.	1590. μ Ci	-0.3%	AMA

Approved By: _____

Date: _____