

## Stored Wipe Data

Biodex Model 086-332 Serial # 08070015 Detector Serial # 9020011

Printed: 05/06/2009 12:03 PM

Print Range: Today

Date	Detector SN	Location	Status	Wide Window		Technologist
				Net CPM	Net Activity	
05/06/09 11:45 AM	9020011	137 Cs Button	PASS	92.0	0.0001 uCi	Nuc Tech
05/06/09 11:44 AM	9020011	Cs137 Rod	PASS	118.0	0.0002 uCi	Nuc Tech
05/06/09 11:40 AM	9020011	Hot Lab	FAIL	12,780.0	37,076 DPM	Nuc Tech
05/06/09 10:23 AM	9020011	camera room	PASS	319.0	925 DPM	Nuc Tech
05/06/09 10:22 AM	9020011	sink	PASS	63.0	183 DPM	Nuc Tech
05/06/09 10:20 AM	9020011	camera 3	FAIL	25,858.0	75,016 DPM	Nuc Tech
05/06/09 10:13 AM	9020011	Desk	PASS	61.0	177 DPM	Nuc Tech
05/06/09 10:09 AM	9020011	Computer Keyboard	PASS	35.0	102 DPM	Nuc Tech
05/06/09 9:49 AM	9020011	Floor	FAIL	48,686.0	141,242 DPM	Nuc Tech

Approved By: \_\_\_\_\_

Date: \_\_\_\_\_