

Wipe Detail Report

Biodex Model 086-332 Serial # 08070015 Detector Serial # 9020011
Printed: 04/27/2009 16:32

Print Range: From 03/01/2009 To 04/27/2009						
04/27/09 16:22	Hot Lab	Restricted, 100%		Time: 60 secs	Status: FAIL	Tech: AMA
	W.W.	Isotope	Background	Net Activity	Net CPM	Trigger
	1	Cs-137	908 DPM	121.044 DPM	41724.0	20.000 DPM
	2	Ga-67	422 DPM	43.994 DPM	22.414.0	20.000 DPM
	3	I-131	153 DPM	19.498 DPM	6.998.0	2.000 DPM
	4	Tc-99m	41 DPM	4.357 DPM	3.369.0	20.000 DPM
		TI-201	131 DPM	8.498 DPM	4.209.0	20.000 DPM
						PASS
04/27/09 12:59	Floor	Restricted, 100%		Time: 60 secs	Status: PASS	Tech: AMA
	W.W.	Isotope	Background	Net Activity	Net CPM	Trigger
	1	Cs-137	908 DPM	17 DPM	6.0	20.000 DPM
	2	Tc-99m	41 DPM	5 DPM	4.0	20.000 DPM
	3	Ga-67	422 DPM	-10 DPM	-5.0	20.000 DPM
	4	TI-201	131 DPM	-14 DPM	-7.0	20.000 DPM
		Cs-137	376 DPM	25 DPM	3.0	20.000 DPM
						PASS
04/27/09 12:49	Hot Lab	Restricted, 100%		Time: 60 secs	Status: PASS	Tech: AMA
	W.W.	Isotope	Background	Net Activity	Net CPM	Trigger
	1	Cs-137	908 DPM	-32 DPM	-11.0	20.000 DPM
	2	Ga-67	422 DPM	-35 DPM	-18.0	20.000 DPM
	3	I-131	153 DPM	0 DPM	0.0	2.000 DPM
	4	Tc-99m	41 DPM	-1 DPM	-1.0	20.000 DPM
		TI-201	131 DPM	-28 DPM	-14.0	20.000 DPM
						PASS
04/27/09 12:42	Hot Lab	Restricted, 100%		Time: 60 secs	Status: FAIL	Tech: AMA
	W.W.	Isotope	Background	Net Activity	Net CPM	Trigger
	1	Cs-137	908 DPM	62.652 DPM	21596.0	20.000 DPM
	2	Ga-67	422 DPM	22.496 DPM	11.461.0	20.000 DPM
	3	I-131	153 DPM	10.214 DPM	3.666.0	2.000 DPM
	4	Tc-99m	41 DPM	2.176 DPM	1.683.0	20.000 DPM
		TI-201	131 DPM	4.373 DPM	2.166.0	20.000 DPM
						PASS
04/02/09 12:19	test	Restricted, 100%		Time: 30 secs	Status: PASS	Tech: AMA
	W.W.	Isotope	Background	Net Activity	Net CPM	Trigger
	1	Cs-137	833 DPM	116 DPM	40.0	20.000 DPM
	2	Tc-99m	41 DPM	0 DPM	0.0	20.000 DPM
	3	Ga-67	357 DPM	71 DPM	36.0	20.000 DPM
		Cs-137	425 DPM	0 DPM	0.0	20.000 DPM
						PASS
04/02/09 11:41	test	Restricted, 100%		Time: 30 secs	Status: PASS	Tech: AMA
	W.W.	Isotope	Background	Net Activity	Net CPM	Trigger
	1	Cs-137	833 DPM	-12 DPM	-4.0	20.000 DPM
	2	Tc-99m	41 DPM	21 DPM	16.0	20.000 DPM
	3	Ga-67	357 DPM	-31 DPM	-16.0	20.000 DPM
		Cs-137	425 DPM	-65 DPM	-8.0	20.000 DPM
						PASS
04/02/09 11:34	test	Restricted, 100%		Time: 30 secs	Status: PASS	Tech: AMA
	W.W.	Isotope	Background	Net Activity	Net CPM	Trigger
	1	Cs-137	833 DPM	41 DPM	14.0	20.000 DPM
	2	Tc-99m	41 DPM	3 DPM	2.0	20.000 DPM
	3	Ga-67	357 DPM	43 DPM	22.0	20.000 DPM
		Ba-133	682 DPM	-81 DPM	-10.0	20.000 DPM
						PASS

Approved By: _____

Date: _____