


Wipe Test Results

Biodex Model 086-332 Serial # 08070015 Detector Serial # 9020011

04/27/2009 16:22
Test Date



Highlighted ROI: Ga-67 (53-424 keV)

Full Spectrum

Net Cnts: 41,781
Total Net CPM: 41,781.0
Test Time (secs): 60
Net Cnts>800keV: 57

Wide Window

Total Efficiency: 34.47%
Background: 908 DPM
Net Activity: 121,044 DPM
Net CPM: 41,724.0
ROI: 25 - 800 keV
Trigger Level: 20,000 DPM

Location: Hot Lab

Technologist: AMA

Status: **FAIL - Wide Window and Isotopes**

	Isotope	Background	Net Activity	Net CPM	Trigger	Status
1	Ga-67	422 DPM	43,994 DPM	22,414.0	20,000 DPM	FAIL
2	I-131	153 DPM	19,498 DPM	6,998.0	2,000 DPM	FAIL
3	Tc-99m	41 DPM	4,357 DPM	3,369.0	20,000 DPM	PASS
4	TI-201	131 DPM	8,498 DPM	4,209.0	20,000 DPM	PASS

Approved By: _____

Date: _____